



SATYENDRA NATH BOSE NATIONAL CENTRE FOR BASIC SCIENCES
[Funded by the Department of Science & Technology, Government of India]
BLOCK JD, SECTOR III, SALT LAKE, KOLKATA- 700 098
PHONE: +91-(0) 33-2335 5706-08, 2335 3057/61, 2335 0312/1313
FAX: +91-(0) 33-2335 3477/1364
EMAIL: santosh@bose.res.in

Ref. SNB/ENQ/MM/Co2 Incubator /14-15/867(WP)

25/09/2014


Sealed quotations is hereby invited for the following items in two parts (Technical and Commercial bids). One large envelope containing two smaller envelopes containing Part A: Technical Bid and Part B: Commercial Bid need to be submitted separately – Two smaller envelopes should be super-scribed "**Technical Bid**" / "**Commercial Bid**" as the case may be.

Sl. No.	PARTICULARS	Qty
	<p>Specification of CO₂ Incubator: Total Volume: about 6.5Cu.ft. / 180 lit. Perforated Shelves (standard): 4 (Sturdy stainless steel shelves.) Polished stainless steel interior with coved corners. Microbiological filters should be provided on gas inlet and sample ports and it should be readily accessible for simplified routine replacement. Inner door gasket should be removable and cleanable, and ensure a tight seal. Temp range 5 °C to above ambient: 50 °C; Temp. Variation at 37 °C: ± 0.3 °C CO₂ range: 0 – 20 % Inlet Pressure: 15PSGI (1.0bar) CO₂ Sensor: T/C Readability & Setability: 0.1% ; Humidity : Ambient to 95% RH at 37°C TC (thermal conductivity) should be uniform and sensitive IR (infrared) CO₂ gas sensors should sense CO₂ properly and should display accurately and contamination control options should be there. Standard remote alarm contacts (NO, NC, COM) should be there.</p>	01
<p>Note: Attached herewith "Annexure I" (as check list) to be duly filled in by the bidder and should be submitted with the Technical Bid only. Offer received without any of the relevant information / certificate / document asked as per "Annexure I" from sl. nos. 01 to 18 may not be considered. The Centre reserves the right to accept or reject offer of the Tenderer. The Centre's decision shall be final and binding on the tenderer. Attached documents should be duly marked.</p>		
<p>Technical bid should contain complete technical information literature/working manual of the quoted item, & authorization certificate of the manufacturer. Price Bid - In case of imported item CIF Kolkata airport price should be mentioned and for indigenous item FOR basis up to S. N. Bose National Centre for Basic Sciences, Kolkata may be quoted.</p>		

Note:

- 1) The quoted price should be inclusive of delivery charges.
- 2) Validity should be for 90 days from the date of opening
- 3) Minimum one (1) year standard onsite warranty is to be provided.
- 4) Our Payment term is subject to after delivery and installation only.
- 5) Delivery period should be mentioned in the quotation.
- 6) Our ref. no. should be mentioned on top of the quotation envelope.
- 7) Quotation for the above item should reach this office by **15th October, 2014**

Thanking you,
Yours faithfully,


S. K. Singh
AR (Purchase)

